

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXD-0001
Location:	GLEN ROCK POLICE	
Calibration File No.:	00588	Calib. Date: 01/06/2017
Certification File No.:	00589	Calib. No.: 00028
Linearity File No.:	00590	Cert. Date: 01/06/2017
Solution File No.:	00597	Cert. No.: 00022
Sequential File No.:	00597	Lin. Date: 01/06/2017
		Lin. No.: 00022
		Soln. Date: 02/15/2017
		Soln. No.: 00133
		File Date: 02/15/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZJ-0091
Solution Control Lot:	16150	Expires: 07/06/2018
		Bottle No.: 1121

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:30S		
Control 1 EC	0.098%	00:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:31S		
Control 2 EC	0.099%	00:32S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:33S		
Control 3 EC	0.098%	00:33S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: MCGOVERN

First Name: J. SCOTT

MI: -

Signature: \_\_\_\_\_



Badge No.: 8130

Date: 02/15/2017