

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXD-0001
Location:	GLEN ROCK POLICE	
Calibration File No.:	00717	Calib. Date: 10/11/2019
Certification File No.:	00718	Cert. No.: 00036
Linearity File No.:	00719	Cert. Date: 10/11/2019
Solution File No.:	00722	Lin. Date: 10/11/2019
Sequential File No.:	00722	Soln. Date: 11/14/2019
		Lin. No.: 00030
		Soln. No.: 00166
		File Date: 11/14/2019

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZJ-0091
Solution Control Lot:	19110	Expires: 04/09/2021
		Bottle No.: 0044

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:49S		
Control 1 EC	0.100%	11:50S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:50S		
Control 2 EC	0.099%	11:51S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:51S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:52S		
Control 3 EC	0.099%	11:52S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:53S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCGOVERN

First Name: J. SCOTT

MI: -

Signature: 

Badge No.: 8130

Date: 11/14/2019