

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXD-0001
Location:	GLEN ROCK POLICE	
Calibration File No.:	00628	Calib. Date: 10/17/2017
Certification File No.:	00629	Calib. No.: 00031
Linearity File No.:	00630	Cert. Date: 10/17/2017
Solution File No.:	00634	Cert. No.: 00025
Sequential File No.:	00634	Lin. Date: 10/17/2017
		Lin. No.: 00025
		Soln. Date: 11/16/2017
		Soln. No.: 00142
		File Date: 11/16/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZJ-0090
Solution Control Lot:	17110	Expires: 03/20/2019
		Bottle No.: 0574

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:30S		
Control 1 EC	0.101%	17:31S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	17:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:32S		
Control 2 EC	0.100%	17:32S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	17:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:33S		
Control 3 EC	0.100%	17:34S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	17:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCGOVERN

First Name: J. SCOTT

MI: -

Signature: 

Badge No.: 8130

Date: 11/16/2017