

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXD-0001
Location:	GLEN ROCK POLICE	
Calibration File No.:	00628	Calib. Date: 10/17/2017
Certification File No.:	00629	Calib. No.: 00031
Linearity File No.:	00630	Cert. Date: 10/17/2017
Solution File No.:	00641	Cert. No.: 00025
Sequential File No.:	00641	Lin. Date: 10/17/2017
		Lin. No.: 00025
		Soln. Date: 01/23/2018
		Soln. No.: 00144
		File Date: 01/23/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZJ-0090
Solution Control Lot:	17110	Expires: 03/20/2019
		Bottle No.: 0573

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:28S		
Control 1 EC	0.101%	15:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:30S		
Control 2 EC	0.101%	15:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:31S		
Control 3 EC	0.100%	15:32S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:33S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCGOVERN

First Name: J. SCOTT

MI: -

Signature:  _____

Badge No.: 8130

Date: 01/23/2018