

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXD-0001
Location:	GLEN ROCK POLICE	
Calibration File No.:	00685	Calib. Date: 11/15/2018
Certification File No.:	00686	Calib. No.: 00034
Linearity File No.:	00687	Cert. Date: 11/15/2018
Solution File No.:	00691	Cert. No.: 00028
Sequential File No.:	00691	Lin. Date: 11/15/2018
		Lin. No.: 00028
		Soln. Date: 12/22/2018
		Soln. No.: 00156
		File Date: 12/22/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZJ-0090
Solution Control Lot:	18090	Expires: 03/13/2020
		Bottle No.: 0376

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:36S		
Control 1 EC	0.100%	18:37S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:37S		
Control 2 EC	0.099%	18:38S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:39S		
Control 3 EC	0.098%	18:39S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	18:39S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:40S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MCGOVERN

First Name: J. SCOTT

MI: -

Signature: _____



Badge No.: MCGOVERN

Date: 12/22/2018