

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXD-0001
Location: GLEN ROCK POLICE
Calibration File No.: 00685 Calib. Date: 11/15/2018 Calib. No.: 00034
Certification File No.: 00686 Cert. Date: 11/15/2018 Cert. No.: 00028
Linearity File No.: 00687 Lin. Date: 11/15/2018 Lin. No.: 00028
Solution File No.: 00696 Soln. Date: 01/22/2019 Soln. No.: 00157
Sequential File No.: 00696 File Date: 01/22/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZJ-0090
Control Solution %: 0.100% Expires: 03/13/2020
Solution Control Lot: 18090 Bottle No.: 0319

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:01S		
Control 1 EC	0.099%	16:02S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:03S		
Control 2 EC	0.099%	16:03S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	16:03S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:04S		
Control 3 EC	0.097%	16:05S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:05S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:05S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCGOVERN

First Name: J. SCOTT

MI: -

Signature: 

Badge No.: 8130

Date: 01/22/2019