

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0001
Location:	GLEN ROCK POLICE		
Calibration File No.:	00685	Calib. Date:	11/15/2018
Certification File No.:	00686	Cert. No.:	00034
Linearity File No.:	00687	Cert. Date:	11/15/2018
Solution File No.:	00700	Lin. Date:	11/15/2018
Sequential File No.:	00700	Lin. No.:	00028
		Soln. Date:	02/25/2019
		Soln. No.:	00158
		File Date:	02/25/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZJ-0090
Solution Control Lot:	18090	Expires:	03/13/2020
		Bottle No.:	0768

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:21S		
Control 1 EC	0.099%	10:22S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:22S		
Control 2 EC	0.099%	10:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:24S		
Control 3 EC	0.098%	10:25S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:25S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCGOVERN

First Name: J. SCOTT

MI: -

Signature: 

Badge No.: 8130

Date: 02/25/2019