

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXD-0001
Location:	GLEN ROCK POLICE	
Calibration File No.:	00670	Calib. Date: 05/31/2018
Certification File No.:	00671	Calib. No.: 00033
Linearity File No.:	00672	Cert. Date: 05/31/2018
Solution File No.:	00675	Cert. No.: 00027
Sequential File No.:	00675	Lin. Date: 05/31/2018
		Lin. No.: 00027
		Soln. Date: 07/06/2018
		Soln. No.: 00150
		File Date: 07/06/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZJ-0091
Solution Control Lot:	18090	Expires: 03/13/2020
		Bottle No.: 0741

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:28D		
Control 1 EC	0.100%	12:29D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:29D		
Control 2 EC	0.098%	12:30D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:31D		
Control 3 EC	0.099%	12:31D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	12:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:32D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCGOVERN

First Name: J. SCOTT

MI: -

Signature: 

Badge No.: 8130

Date: 07/06/2018