

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARXD-0001 |
| Location: | GLEN ROCK POLICE | |
| Calibration File No.: | 00670 | Calib. Date: 05/31/2018 |
| Certification File No.: | 00671 | Cert. No.: 00033 |
| Linearity File No.: | 00672 | Cert. Date: 05/31/2018 |
| Solution File No.: | 00676 | Cert. No.: 00027 |
| Sequential File No.: | 00676 | Lin. Date: 05/31/2018 |
| | | Lin. No.: 00027 |
| | | Soln. Date: 07/24/2018 |
| | | Soln. No.: 00151 |
| | | File Date: 07/24/2018 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDZJ-0091 |
| Solution Control Lot: | 18090 | Expires: 03/13/2020 |
| | | Bottle No.: 0745 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 09:25D | | |
| Control 1 EC | 0.100% | 09:25D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 09:25D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:26D | | |
| Control 2 EC | 0.099% | 09:27D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 09:27D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:27D | | |
| Control 3 EC | 0.099% | 09:28D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 09:28D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 09:29D | | |

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCGOVERN

First Name: J. SCOTT

MI: -

Signature: 

Badge No.: 8130

Date: 07/24/2018