

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXD-0001
Location:	GLEN ROCK POLICE		
Calibration File No.:	00670	Calib. Date:	05/31/2018
Certification File No.:	00671	Cert. No.:	00033
Linearity File No.:	00672	Cert. Date:	05/31/2018
Solution File No.:	00678	Lin. Date:	05/31/2018
Sequential File No.:	00678	Lin. No.:	00027
		Soln. Date:	08/28/2018
		Soln. No.:	00152
		File Date:	08/28/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDZJ-0091
Solution Control Lot:	18090	Expires:	03/13/2020
		Bottle No.:	0378

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:57D		
Control 1 EC	0.099%	09:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:58D		
Control 2 EC	0.097%	09:59D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:59D		
Control 3 EC	0.098%	10:00D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCGOVERN

First Name: J. SCOTT

MI: -

Signature: 

Badge No.: 8130

Date: 08/28/2018