

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXD-0001  
**Location:** GLEN ROCK POLICE  
**Calibration File No.:** 00670 **Calib. Date:** 05/31/2018 **Calib. No.:** 00033  
**Certification File No.:** 00671 **Cert. Date:** 05/31/2018 **Cert. No.:** 00027  
**Linearity File No.:** 00672 **Lin. Date:** 05/31/2018 **Lin. No.:** 00027  
**Solution File No.:** 00684 **Soln. Date:** 11/13/2018 **Soln. No.:** 00154  
**Sequential File No.:** 00684 **File Date:** 11/13/2018

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDZJ-0091  
**Control Solution %:** 0.100% **Expires:** 03/13/2020  
**Solution Control Lot:** 18090 **Bottle No.:** 0208

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:41S		
Control 1 EC	0.101%	18:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	18:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:43S		
Control 2 EC	0.100%	18:43S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	18:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:44S		
Control 3 EC	0.099%	18:45S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:46S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: MCGOVERN

First Name: J. SCOTT

MI: -

Signature: \_\_\_\_\_

Badge No.: 8130

Date: 11/13/2018