

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXD-0001  
Location: GLEN ROCK POLICE  
Calibration File No.: 00569 Calib. Date: 07/18/2016 Calib. No.: 00027  
Certification File No.: 00570 Cert. Date: 07/18/2016 Cert. No.: 00021  
Linearity File No.: 00571 Lin. Date: 07/18/2016 Lin. No.: 00021  
Solution File No.: 00574 Soln. Date: 08/15/2016 Soln. No.: 00127  
Sequential File No.: 00574 File Date: 08/15/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZJ-0090  
Control Solution %: 0.100% Expires: 04/13/2017  
Solution Control Lot: 15D138 Bottle No.: 1315

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:14D		
Control 1 EC	0.099%	10:15D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:15D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:15D		
Control 2 EC	0.098%	10:16D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:16D		
Control 3 EC	0.098%	10:17D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: MCGOVERN

First Name: J. SCOTT

MI: -

Signature: 

Badge No.: 8130

Date: 08/15/2016