

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARXD-0001 |
| Location: | GLEN ROCK POLICE | |
| Calibration File No.: | 00569 | Calib. Date: 07/18/2016 |
| Certification File No.: | 00570 | Calib. No.: 00027 |
| Linearity File No.: | 00571 | Cert. Date: 07/18/2016 |
| Solution File No.: | 00576 | Cert. No.: 00021 |
| Sequential File No.: | 00576 | Lin. Date: 07/18/2016 |
| | | Lin. No.: 00021 |
| | | Soln. Date: 09/22/2016 |
| | | Soln. No.: 00128 |
| | | File Date: 09/22/2016 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDZJ-0090 |
| Solution Control Lot: | 16150 | Expires: 07/06/2018 |
| | | Bottle No.: 1415 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 13:12D | | |
| Control 1 EC | 0.099% | 13:13D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 13:13D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 13:14D | | |
| Control 2 EC | 0.099% | 13:14D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 13:14D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 13:15D | | |
| Control 3 EC | 0.099% | 13:16D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 13:16D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 13:16D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MCGOVERN

First Name: J. SCOTT

MI: -

Signature: _____



Badge No.: 8130

Date: 09/22/2016