

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXD-0001  
Location: GLEN ROCK POLICE  
Calibration File No.: 00569 Calib. Date: 07/18/2016 Calib. No.: 00027  
Certification File No.: 00570 Cert. Date: 07/18/2016 Cert. No.: 00021  
Linearity File No.: 00571 Lin. Date: 07/18/2016 Lin. No.: 00021  
Solution File No.: 00578 Soln. Date: 10/22/2016 Soln. No.: 00129  
Sequential File No.: 00578 File Date: 10/22/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZJ-0090  
Control Solution %: 0.100% Expires: 07/06/2018  
Solution Control Lot: 16150 Bottle No.: 1464

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:26D		
Control 1 EC	0.100%	17:27D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:28D		
Control 2 EC	0.100%	17:28D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	17:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:29D		
Control 3 EC	0.099%	17:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:30D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: MCGOVERN

First Name: J. SCOTT

MI: -

Signature: 

Badge No.: 8130

Date: 10/22/2016