

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXD-0001
Location: GLEN ROCK POLICE
Calibration File No.: 00569 Calib. Date: 07/18/2016 Calib. No.: 00027
Certification File No.: 00570 Cert. Date: 07/18/2016 Cert. No.: 00021
Linearity File No.: 00571 Lin. Date: 07/18/2016 Lin. No.: 00021
Solution File No.: 00581 Soln. Date: 11/14/2016 Soln. No.: 00130
Sequential File No.: 00581 File Date: 11/14/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZJ-0090
Control Solution %: 0.100% Expires: 07/06/2018
Solution Control Lot: 16150 Bottle No.: 0865

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:38S		
Control 1 EC	0.098%	21:39S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:39S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:40S		
Control 2 EC	0.098%	21:40S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:41S		
Control 3 EC	0.098%	21:42S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:42S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:42S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCGOVERN

First Name: J. SCOTT

MI: -

Signature: 

Badge No.: 8130

Date: 11/14/2016