

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXD-0001
Location:	GLEN ROCK POLICE	
Calibration File No.:	00569	Calib. Date: 07/18/2016
Certification File No.:	00570	Calib. No.: 00027
Linearity File No.:	00571	Cert. Date: 07/18/2016
Solution File No.:	00583	Cert. No.: 00021
Sequential File No.:	00583	Lin. Date: 07/18/2016
		Lin. No.: 00021
		Soln. Date: 12/12/2016
		Soln. No.: 00131
		File Date: 12/12/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDZJ-0090
Solution Control Lot:	16150	Expires: 07/06/2018
		Bottle No.: 0975

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:40S		
Control 1 EC	0.101%	16:41S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:42S		
Control 2 EC	0.100%	16:42S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:42S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:43S		
Control 3 EC	0.100%	16:44S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	16:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:44S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MCGOVERN

First Name: J. SCOTT

MI: -

Signature: _____



Badge No.: 8130

Date: 12/12/2016