



STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF CONSUMER AFFAIRS
 OFFICE OF WEIGHTS AND MEASURES
 1261 ROUTES 1 & 9 SOUTH
 AVENEL, NEW JERSEY 07001



Date: November 12, 2010

GLEN ROCK POLICE DEPT.
 1 Harding Plaza
 Glen Rock, NJ 07452

Subject: TEST REPORT - NJ110275 - LASER SPEED GUN DISTANCE CALIBRATION COURSE

TEST OF ACCURACY ONLY

The course(s) which you submitted to the New Jersey State Metrology Laboratory were/was compared with the State Standard (NIST Test Number 821/263710-00) on this date.

<u>DESIGNATION</u>	<u>ACTUAL</u>	<u>ID MARKS</u>
CENTER OF BASE POST TO:		GRPD-LC
CENTER OF FIRST TARGET (1)	50 FEET (+/- <3/16")	
CENTER OF SECOND TARGET (2)	75 FEET (+/- <3/16")	

This certificate shall not be reproduced except in full without the written permission of the State Superintendent of Weights and Measures.

Robert Campanelli
 Acting State Superintendent

